

# BARD® SPECIALS MATRIX



Product	4 French	5 French	6 French	7 French	Notes
<b>Woven Catheters</b> <ul style="list-style-type: none"> <li>• Number of Poles</li> <li>• Usable Length (5cm increments)</li> <li>• Electrode Configuration                             <ul style="list-style-type: none"> <li>– Minimum Spacing Increments</li> <li>– Proximal Band Width</li> <li>– Distal Tip Length</li> </ul> </li> </ul>	Up to 4 50-110cm	Up to 6 50-150cm	Up to 24 50-150cm	Up to 12 50-150cm	Electrodes must be placed within 42cm of tip (19cm for High Torque Shaft). High Torque Shaft available in 6 French.
<b>WovenFlexie™ Catheters</b> <ul style="list-style-type: none"> <li>• Number of Poles</li> <li>• Usable Length (5cm increments)</li> <li>• Electrode Configuration                             <ul style="list-style-type: none"> <li>– Minimum Spacing Increments</li> <li>– Proximal Band Width</li> <li>– Distal Tip Length</li> </ul> </li> </ul>	Not Available	Up to 6 50-140cm	Up to 10 50-140cm	Not Available	Electrodes must be placed within 42cm of the tip.
<b>Viking™ Catheters</b> <ul style="list-style-type: none"> <li>• Number of Poles</li> <li>• Usable Length (5cm increments)</li> <li>• Electrode Configuration                             <ul style="list-style-type: none"> <li>– Minimum Spacing Increments</li> <li>– Proximal Band Width</li> <li>– Distal Tip Length</li> </ul> </li> </ul>	Not Available	Up to 10 50-135cm	Up to 10 50-135cm	Not Available	Electrodes must be placed within 13.5cm of the tip.
<b>Tango™ 5 FR Catheter &amp; Viking SoftTip™ 6 FR Catheter</b> <ul style="list-style-type: none"> <li>• Number of Poles</li> <li>• Usable Length (5cm increments)</li> <li>• Electrode Configuration                             <ul style="list-style-type: none"> <li>– Minimum Spacing Increments</li> <li>– Proximal Band Width</li> <li>– Distal Tip Length</li> </ul> </li> </ul>	Not Available	Up to 10 65-135cm	Up to 10 65-135cm	Not Available	Electrodes must be placed within 6.5cm of the tip.
<b>EP•XT® Catheters</b> Uni-Directional Rotating Actuator Handle	Not Available	Not Available	Up to 10 110cm	Not Available	Minimum spacing 2mm between distal tip electrode and second electrode and 1mm thereafter. Electrodes must be placed 2cm distal to Tip Joint (where the tip/shaft transition occurs).
<ul style="list-style-type: none"> <li>• Number of Poles</li> <li>• Usable Length</li> <li>• Electrode Configuration                             <ul style="list-style-type: none"> <li>– Minimum Spacing Increments</li> <li>– Proximal Band Width</li> <li>– Distal Tip Length</li> </ul> </li> <li>• Curve Size (determines tip section length)</li> </ul>			See notes 1mm 2mm Large (4cm tip)		
<b>Dynamic XT™ Catheter and Dynamic Tip™ Catheter</b> Uni-Directional Push/Pull Actuator Handle	Not Available	Not Available	Up to 10 60-110cm	Not Available	Minimum spacing 2mm between distal tip electrode and second electrode and 1mm thereafter. Electrodes must be placed 2cm distal to Tip Joint (where the tip/shaft transition occurs).
<ul style="list-style-type: none"> <li>• Number of Poles</li> <li>• Usable Length (5cm increments)</li> <li>• Electrode Configuration                             <ul style="list-style-type: none"> <li>– Minimum Spacing Increments</li> <li>– Proximal Band Width</li> <li>– Distal Tip Length</li> </ul> </li> <li>• Curve Size (determines tip section length)</li> </ul>			See notes 1mm 2mm Large (4cm tip)		
<b>Orbiter ST™ Catheters</b> Bi-Directional Push/Pull Actuator Handle	Not Available	Not Available	Not Available	4 to 20† 95cm	Minimum spacing is 1/3/1mm. No Electrodes allowed within 8mm of Tip Joint (where the tip/shaft transition occurs).
<ul style="list-style-type: none"> <li>• Number of Poles</li> <li>• Usable Length</li> <li>• Electrode Configuration                             <ul style="list-style-type: none"> <li>– Minimum Spacing Increments</li> <li>– Proximal Band Width</li> <li>– Distal Tip Length</li> </ul> </li> <li>• Curve Size (determines tip section length)</li> </ul>				See notes 1mm 2mm Large (187mm tip) Small (157mm tip) Custom (62-187mm tip, 1mm increments)	

Contact your Territory Manager for Connector Type options. Electrode spacing is measured from edge-to-edge unless otherwise specified.

\* Not available in all geographies. Contact your Territory Manager for product availability. † Electrode Quantity may be limited by Tip Length.

